



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/613,335

07/03/2003

Walter A. Zohmann

10012.7

5090

21999 7590 10/16/2008

KIRTON AND MCCONKIE
60 EAST SOUTH TEMPLE,
SUITE 1800
SALT LAKE CITY, UT 84111

EXAMINER

CAMPBELL, VICTORIA P

ART UNIT

PAPER NUMBER

3763

MAIL DATE

DELIVERY MODE

10/16/2008

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/613,335	Applicant(s) ZOHMANN, WALTER A.	
	Examiner VICTORIA P. CAMPBELL	Art Unit 3763	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 07 July 2008.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-9 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-9 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

This is the third Office Action following the Request for Continued Examination based on the 10/613335 application filed July 3, 2003. Claims 1-9 as presented are currently pending and considered below.

Claim Rejections - 35 USC § 112

1. The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.
2. Claims 7-9 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.
3. Claims 7-9 recite the limitation "said fenestrated needle" in line 6 of claim 7, line 1 of claim 8, and lines 2-3 of claim 9. There is insufficient antecedent basis for this limitation in the claims.

Claim Rejections - 35 USC § 103

4. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Art Unit: 3763

5. The factual inquiries set forth in *Graham v. John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

1. Determining the scope and contents of the prior art.
2. Ascertaining the differences between the prior art and the claims at issue.
3. Resolving the level of ordinary skill in the pertinent art.
4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

6. Claims 1, 2, and 7-9 are rejected under 35 U.S.C. 103(a) as being unpatentable over USPN 5,250,035 to Smith et al in view of USPGPub 2002/0123723 to Sorenson et al.

Regarding the above claims, Smith et al teach a hollow needle (28), a needle hub (32) having a hollow interior (38) at the proximate end of the hollow needle (34), and a stylet cap (60) on a proximal end of a stylet (68), the stylet being freely movable within the needle (Col. 4, lines 43-45), wherein the stylet cap creates a releasably secure fit with the needle hub (66 received into 38). Smith et al also teach the process of identifying the dermal area of the patient (Col. 3, lines 27-29), inserting and advancing the needle into the dermal area (Col. 4, lines 63-67), withdrawing the stylet (Col. 5, lines 6-8), and injecting an anesthetic (Col. 4, lines 58-60), wherein the needle further comprises a needle hub (32), and wherein withdrawal of the stylet comprising observing a backflow of fluid (Col. 5, lines 8-15).

Smith et al fail to teach or disclose the hollow needle having at least 3 fenestrations longitudinally disposed along alternate sides of the needle. However,

Art Unit: 3763

Sorenson et al teach a plurality of fenestrations (85, of which there are at least 3) disposed on alternate sides of the needle (Fig. 1).

Smith et al and Sorenson et al are analogous art because they are from the same field of endeavor/problem solving area of medical needles. At the time of the invention, it would have been obvious to one of ordinary skill in the art, having the teachings of Smith et al and Sorenson et al before him or her to modify the needle of Smith et al to include the multiple apertures of Sorenson et al because doing so provides a wider distribution of fluid than with a single opening (Sorenson et al, Paragraph [0033]). Therefore, it would have been obvious to combine Smith et al with Sorenson et al to obtain the invention in the instant claims.

7. Claims 3-6 are rejected under 35 U.S.C. 103(a) as being unpatentable over Smith et al and Sorenson et al in further view of USPGPub 2002/055715 to Young et al.

Regarding claims 3 and 4, Smith et al and Sorenson et al disclose the invention of claims 1 and 2 as described above, but fail to teach or disclose a fenestration indicator or a magnifying window on the needle hub. However, Young et al teach a needle hub (10) containing both a fenestration indicator (16) and a magnifier (17).

Regarding claims 5 and 6, Smith et al teach a hollow needle (28) being bounded by an occluded tip (72), a needle hub (32) at the proximate end of the hollow needle (34), and a stylet cap (60) on a proximal end of a stylet (68), wherein the stylet cap creates a releasably secure fit with the needle hub (66 received into 38), the stylet being freely movable within the needle (Col. 4, lines 43-45), wherein the stylet occludes the fenestrations (Fig. 11).

Smith et al fail to teach the needle having at least three longitudinally disposed fenestrations. Smith et al also fail to teach the needle hub having at least one fenestration indicator and a magnifying window.

Sorenson et al teach a plurality of fenestrations (85) disposed on alternate sides of the needle (Fig. 1). Combination of Smith et al and Sorenson et al is reasoned above.

Young et al teach a needle hub (10) containing both a fenestration indicator (16) and a magnifier (17).

Smith et al, Sorenson et al, and Young et al are analogous art because they are from the same field of endeavor/problem solving area of medical needles. At the time of the invention, it would have been obvious to one of ordinary skill in the art, having the teachings of Smith et al, Sorenson et al, and Young et al before him or her to modify the needle system of Smith et al and Sorenson et al to include the fenestration indicator and magnifying window of Young et al because the indicator allows the user to properly orient the hub during low light conditions (Young et al, Paragraph [0031]) and the magnifier decreases the recognition time from when the fluid first enters the hub (Young et al, Abstract). Therefore, it would have been obvious to combine Smith et al and Sorenson et al with Young et al to obtain the invention in the instant claims.

Response to Arguments

8. Applicant's arguments filed July 7, 2008 have been fully considered but they are not persuasive.

Art Unit: 3763

9. In response to applicant's argument that the references fail to show certain features of applicant's invention, it is noted that the features upon which applicant relies (i.e., the location of the fenestrations within 1-2 mm and most preferably within 0.17 inches of each other) are not recited in the rejected claim(s). Although the claims are interpreted in light of the specification, limitations from the specification are not read into the claims. See *In re Van Geuns*, 988 F.2d 1181, 26 USPQ2d 1057 (Fed. Cir. 1993).

10. Additionally, the examiner would like to note that the claims as presented call for "at least three fenestrations [...] isolated on a distal end of said hollow needle" (claim 1, similar to limitations presented in claims 5 and 7) which is exactly the structure presented by Sorenson et al. The fenestrations (85) do not occur along the entire length of the needle (30), just a distal portion of the needle, as seen in Fig. 1.

Conclusion

11. **THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of

Art Unit: 3763

the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action

Any inquiry concerning this communication or earlier communications from the examiner should be directed to VICTORIA P. CAMPBELL whose telephone number is (571)270-5035. The examiner can normally be reached on Monday-Thursday, 7-5:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nicholas Lucchesi can be reached on 571-272-4977. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Victoria P Campbell
Examiner, AU 3763

/Nicholas D Lucchesi/
Supervisory Patent Examiner, Art Unit 3763